

## Do we need to enforce the homogeneous Neumann condition on the torso for solving the inverse electrocardiographic problem?

Judit Chamorro-Servent, Laura Bear, Josselin Duchateau, Mark Potse, Rémi Dubois, Yves Coudière

#### ▶ To cite this version:

Judit Chamorro-Servent, Laura Bear, Josselin Duchateau, Mark Potse, Rémi Dubois, et al.. Do we need to enforce the homogeneous Neumann condition on the torso for solving the inverse electrocardiographic problem?. Computing in Cardiology, Sep 2016, Vancouver, Canada. 2016. hal-01378822

#### HAL Id: hal-01378822 https://inria.hal.science/hal-01378822v1

Submitted on 10 Oct 2016

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# Do we need to enforce the homogeneous Neumann condition on the torso for solving the inverse electrocardiographic problem?

<u>Judit Chamorro-Servent</u><sup>1,2,3</sup>, Laura Bear<sup>1,4,5</sup>, Josselin Duchateau<sup>1,4,5</sup>, Mark Potse<sup>1,2,3</sup>, Rémi Dubois<sup>1,4,5</sup>, Yves Coudière<sup>1,2,3</sup>









achieved by  $\lambda = 0$  (MFS without Neumann

conditions) and  $\lambda = 0.5$  (Standard MFS).




<sup>1</sup>IHU Liryc, Electrophysiology and Heart Modeling Institute, foundation Bordeaux Université, Pessac-Bordeaux, France

<sup>2</sup>CARMEN Research Team, INRIA, Bordeaux, France

<sup>3</sup>Univ. Bordeaux, IMB, UMR 5251, CNRS, INP-Bordeaux, Talence, France <sup>4</sup>Univ. Bordeaux, CRCTB, U1045, Bordeaux, France <sup>5</sup>INSERM, CRCTB, UMR5251, Bordeaux, France

#### Introduction

➤ Robust calculations of the inverse electrocardiographic problem may require accurate specification of boundary conditions at the torso and cardiac surfaces [1].




 $-div(\phi)$ , in  $\Omega$ . Under: Dirichlet conditions:  $\phi = \phi_T$  on  $\Gamma_T$  (1) Homogeneous Neumann conditions:  $\partial_n \phi = 0$  on  $\Gamma_T$ 

The aim of the inverse problem is to find  $\phi = \phi_F$  on  $\Gamma_F$ 

Geometrical meshes used: The body is cut at the top and bottom of the torso and the arms.

The classical formulation of the ECGI inverse problem with the method of fundamental solution (MFS) involves a linear system[2]. That can be split into two submatrices:



- We can remark that:
- The contribution of the submatrix  $B_1$  and  $B_0$  in the whole solution is related to their respectives norms.
- Due to the meshing of the torso, non-phisiological conditions are applied to the top, bottom of the torso and the arms.
- Accurate computation of the normals at torso surface are required.
- These remarks drove us to the question: Do we need to enforce the homogeneous Neumann conditions in the MFS inverse problem?
- > To solve this question:
- We introduced a weighting coefficient in the cost function (*Method 1*), so as to change the balance between the contributions of the Neumann and Dirichlet conditions.
- We proposed to consider the Neumann conditions as a constraint (*Method 2*), so that it is resolved exactly.

#### Methods

#### 1. Minimization with weighed boundary conditions

To evaluate the contribution of both boundary conditions:

$$J_{\lambda}(a) = \frac{1}{2} \left( (1 - \lambda)^{2} \|B_{0}a - \phi_{T}\|_{2}^{2} + \lambda \|B_{1}a\|_{2}^{2} + \alpha \|a\|_{2}^{2} \right), \ \lambda \in [0, 1]$$
 (4)

The solution of (4) was found by solving the following system using Tikhonov regularization

$$\begin{pmatrix} (1-\lambda)B_0 \\ \lambda B_1 \end{pmatrix} a = \begin{pmatrix} \lambda \phi_T \\ 0 \end{pmatrix}$$
 (5)

- Choosing:
  - $\lambda$  = 0.5 corresponds to standard MFS.
  - $\lambda = 0$  corresponds to MFS without Neumann conditions.
  - $\lambda = \frac{1}{1+\varepsilon}$ , where  $\varepsilon = \|B_1\|/\|B_0\|$  provides a balance between both submatrices.

#### 2. Minimization with the Neumann conditions as constraint

• The Neumann conditions can be introduced as a homogeneous equality constraint. In this case, we look for  $a \in \mathbb{R}^{1+N_s}$  that minimizes

$$J_{0}(a) = \frac{1}{2} (\|B_{0}a - \phi_{T}\|_{2}^{2} + \alpha \|a\|_{2}^{2}), \text{s.t. } B_{1}a = 0 \text{ in } \mathbb{R}^{N_{T}}$$
 (6)

This is a saddle-point problem and we can introduce its Lagrangian

$$L(a,\mu) = \frac{1}{2} \|B_0 a - \phi_T\|_2^2 + \frac{1}{2} \alpha \|a\|_2^2 + \mu^T B_1 a, \ \mu \in \mathbb{R}^{N_T}$$
 (7)

The solution is found by solving the block linear system

$$\begin{pmatrix}
B_0^T B_0 + \alpha I d & B_1^T \\
B_1 & 0
\end{pmatrix}
\begin{pmatrix}
\alpha \\
\mu
\end{pmatrix} = \begin{pmatrix}
B_0^T \phi_T \\
0
\end{pmatrix}$$
(8)

We solve this problem with SDPT3, a primal-dual path-following algorithm available on CVX package for convex optimization [3].

### Comparisons

- Five **activation patterns** were simulated : *1 single site pacing* (LV lateral endocardial lasts 500ms), and *4 single spiral waves* with different conductivity coefficients (lasts 3000ms). Simulations were conducted using [4]:
- i. The propagation of electrical activity in a membrane-based **realistic-geometry** computer model, using the governing **monodomain reaction-diffusion equation**. The simulations provide both the theoretical, in-silico  $\phi_T$  and  $\phi_E$  every 1ms.
- ii. The torso model used had **heterogeneous conductivity**.
- Correlation coefficients (CC) and relative errors (RE) for activation times and electrograms were calculated for solutions of *method 1* with  $\lambda = 0$ ,  $\lambda = 0.5$ ,  $\lambda = \frac{1}{1+\varepsilon}$  where  $\varepsilon = \|B_1\|/\|B_0\|$ , and *method 2*.

#### Results

- Respective contribution of B0 and B1
  - The  $\varepsilon = \|B_1\|/\|B_0\|$  depends only on the geometry and location of sources (for which we used a fixed rule defined in [2]). We found  $\varepsilon = 0.0013$ , indicating predominance of  $B_0$  compared to  $B_1$  in the standard MFS.
  - Choosing  $\lambda = 1/(1+\varepsilon) = 0.9987$  in equation (4) leads to a balance between the contribution of the two submatrices.
- Comparison between the reconstructions




Figure 1. Activation maps for electrograms, single site pacing.

CC=0.33

RE=0,44

CC = 0.83

RE=0.51

COMPARISON OF ELECTROGRAMS						
	CC			RE		
	$Q_1$	Median	$Q_3$	$Q_1$	Median	$Q_3$
$\lambda = 0$	0,999	0,999	0,999	0,020	0,034	0,054
$\lambda = 0.5$	0,989	0,995	0,998	0,071	0,107	0,152
$\lambda = 0.9987$	0,258	0,435	0,58	0,997	0,999	1,001
Constrained	0,472	0,688	0,803	0,762	0,981	1,358

Table 2. CC and RE for electrograms, five simulations.

#### **Conclusions**

- ➤ The results of MFS without Neumann condition are in better agreement with the original data. Furthermore, removing the Neumann conditions reduces the forward and inverse computational cost, since the size of the linear system is divided by 2 and the normal vectors do not need to be computed. The ill-posedness of the problem is reduced and consequently the problem is less sensible to the regularization.
- The Lagrange multipliers of *method 2* were significantly higher on the electrodes on the front of the torso than on the ones on the back. This seems to indicate that partial or weighed Neumann conditions (depending on the position of the electrodes) can be explored further in order to refine the inverse problem and improve the resolution of the ECGI problem.
- Note that the results presented here only apply to the MFS method. However, preliminary work seems indicate that homogeneous Neumann conditions at the torso surface have a positive impact on Boundary element method.
- Finally, we used only 5 ventricular datasets obtained by a numerical model. The case of complex atrial activations, and clinical data must be considered for a complete evaluation of the techniques proposed.

#### Acknowledgements

This study received financial support from the French Government as part of the « Investissement d'avenir » program managed by the National Research Agency (ANR), Grant reference ANR-10-IAHU-04 from the Conseil Régional Aquitaine as part of the project « Assimilation de données en cancérologie et cardiologie ».

## References

[1] Barr R, Ramsey M, Spach MS. Relating epicardial to body surface potential distributions by means of transfer

model with parallel processing. IEEE Trans Biomed Eng 2004; 51(8): 1319-1329.

- coefficients based on geometry measurements. IEEE Trans Biomed Eng Jan 1977; BME-24 (1): 1-11.
  [2] Wang Y, Rudy Y. Application of the method of fundamental solutions to potential-based inverse electrocardiography.
- Annals of Biomedical Engineering Aug 2006; 34(8): 1272-1288. [3] Grant M, Boyd S. CVX: Matlab software for disciplined convex programming, version 2.1. <a href="http://cvxr.com/cvx">http://cvxr.com/cvx</a>, March 2014.

[4] Trudel MC, Dubé B, Potse M, Gulrajani RM, Leon LJ. Simulation of propagation in a membrane-based computer heart