

Methodological Design for Integration of Human EEG Data with Behavioral Analyses into Human-Human/Robot Interactions in a Real-World Context

Maria Rodalyn V. Sanchez¹, Satoru Mishima¹, Masayuki Fujiwara², Guangyi Ai³,
Mélanie Jouaiti⁶, Yuliia Kobryn^{6,7}, Sébastien Rimbert⁶, Laurent Bougrain⁶, Patrick Hénaff⁶,
Hiroaki Wagatsuma^{1, 4, 5}

¹Graduate School of Life Science and Systems Engineering
Kyushu Institute of Technology (Kyutech),
2-4 Hibikino, Wakamatsu-ku, Kitakyushu, Fukuoka, Japan
{ sanchez.maria-rodalyn372; q104113s }@mail.kyutech.jp, waga@brain.kyutech.ac.jp

²School of Knowledge Science, Japan Advanced Institute of Science and Technology (JAIST)
Nomi, Ishikawa 923-1211, Japan
m-fujiw@jaist.ac.jp

³Department of Computer Science and Technology, Neusoft Institute Guangdong, Foshan, Guangdong
Province, China
aiguangyi@nuit.edu.cn

⁴RIKEN Center for Brain Science (RIKEN CBS), 2-1 Hirosawa, Wako, Saitama 351-0106, Japan

⁵Artificial Intelligence Research Center, AIST, 2-3-26 Aomi, Koto-Ku, Tokyo 135-0064, Japan

⁶Université de Lorraine, CNRS, Inria, LORIA, F-54000 Nancy, France
{ melanie.jouaiti; laurent.bougrain; patrick.henaff }@loria.fr, { yuliia.kobryn; sebastien.rimbert }@inria.fr

⁷National Technical University of Ukraine “Igor Sikorsky Kyiv Polytechnic Institute”, Kyiv, Ukraine

ABSTRACT. *Analysis of human activities is a complex task that needs multifactorial considerations. So an electroencephalographic (EEG) data analysis can be improved by a conjunction of devices that monitor time courses of multiple types of physiological factors of the subject and counterparts when interactions are on-going in the environment. In this article, we proposed a method to provide the complementary hardware and software treatment that associate devices to be able to synchronize simultaneous data recording to fit the high sampling rate of the EEG signal, such as more than 512 Hz. This method of synchronizing physiological data gathered from three different devices through the use of trigger signals is crucial for an accurate post-analysis and was validated in the experiment. The proposed method is widely applicable in various cases accompanied with EEG measurements and offer a wide possibility in device developments for rehabilitation and communications.*

Keywords: Electroencephalography, Eye-tracker, Motion capture system, Human-Robot interaction, Synchronization

1. Introduction. Studies of human behavior and cognition use questionnaire investigations. However, these contain subjective assessment and need to be completed by physical and physiological measurements of the human behaviors to validate hypotheses such as in psychology [1]. Especially in the interaction between humans or between humans and machines, reproducibility, verifiability, and generality as the foundation of science are crucial not only for the validation of the hypothesis in the scientific question, but also for the development of applications in various fields. In medical field applications, specifically

in rehabilitation, brain-computer interface (BCI) for severely disabled persons, assistive devices for helping a physical movement at the right moment, and communication tools to know what is happening and changing in the interaction [2], obtaining physiological data for analysis is important. Electroencephalography (EEG), which measures the electrical activity of the brain, is suitable for these purposes. It provides a possibility to know what kind of brain processes are on-going in a specific task condition. Measurement and analysis of brain signals are beneficial for providing an insight on how human brain functions work associated with behaviors, actions and intentions.

For the past years, various EEG experiments have been examined such as playing musical instruments [3], performing sports [4, 5], and doing everyday tasks such as walking [6]. Integrating eye-tracking data with EEG recordings have also been done in various studies [7-9], which require effective noise removal methods for EEG data [10, 11]. In the recent years, integrating more types of devices with EEG recording is being explored. However, this poses a challenge of ensuring the synchronization of data gathered from the different devices to guarantee an accurate analysis, since each device has a different internal clock and record data at different sampling rates [12]. In a previous study [13], data synchronization was done through playing a loud sound that was recorded in the eye-tracker audio data, and appears in the EEG data because of the subject's response to the sound. Another study [14], used a software for data acquisition, called lab streaming layer (LSL) [14]. Various software for data synchronization were introduced [12], however, these techniques have the risk for software synchronization delays. To eliminate this risk, hardware implementations for synchronizations through the use of trigger signals [14], are examined, and is what is implemented in this article.

In this article, a method for simultaneously recording EEG with eye-tracker and motion capture system is proposed, together with the corresponding hardware synchronization tool and software treatment for the post-processing of the trigger signals. The three devices chosen are used for the testing and experiment in this article, but implementation is not limited to the specific equipment mentioned, since the methodological design can be applied to different devices in the goal of making reproducible experiments possible.

The objective of this article is to provide a methodological design for integrating EEG with other physiological data recording devices, thus enabling simultaneous recording while ensuring synchronization of the multiple data obtained. In order to achieve this, a generalized system of the methodological design was made to realize the possibility of simultaneously recording EEG together with other devices, including, but not limited to, eye-tracker and motion capture system. Simultaneously recording different devices involves the creation of a data synchronization tool in the form of trigger signals, wherein each device has its own unique trigger signal. In simultaneous recording, all trigger signals are sending at the same time, making the recorded trigger signal the sum of the individual device's unique trigger signals. Results indicate that with the trigger signal data obtained after simultaneous recording, accurate signal reconstruction for the individual device's trigger signal necessary for multiple data synchronization is possible and was demonstrated. The established system for the integration of EEG data with other essential data for the corresponding behavioral analyses is useful in BCI for rehabilitation, human-human and human-robot interaction applications, as presented later in this article.

2. Problem Statement and Preliminaries. The popularization of EEG experiments in various task conditions presents the same challenge concerning the reproducibility in

psychology studies [1]. Independently of the measurement equipment, for a consistent quality of the specification such as sampling rate, the same result should be obtained in a given experimental condition. Alternatively, how much variability exists in the uncontrollable variables such as condition of the human subject, even in the fully-controlled condition is a helpful scientific question. To answer this question and realize the reproducible and replaceable experimental scheme, creation of a methodological design and solving technical problems are necessary.

2.1 Methodological design. As illustrated in Figure 1, one subject's EEG, gaze, and motion data can be simultaneously recorded for examining the human brain activity and behaviors. Moreover, the proposed methodological design provides the possibility for a subject to interact with different counterparts: (1) another human, whose EEG, gaze, and motion data can also be obtained, (2) a robot which runs on a computational model, or (3) a screen where the computational model is embedded in the computer. With this, if the experimenter aims to study a natural human-human interaction, the scheme allows for replacing a human counterpart into a computational model embedded in the computer or robot to verify whether the same phenomenon can be observed or not. If it is a reproducible scheme, the minimization of the complexity of embedded functions in the model reveals the core principle to reproduce the same phenomenon, and then the core can be reproduced in the engineering way.




FIGURE 1. An overview of the methodological design of simultaneous recording and possible counterparts in examining human-human/robot interactions.

2.2 Technical Challenges. In the aim of integrating EEG with other monitoring devices, such as eye-tracking device and motion capture system, we examined a possible way to synchronize every data in a consistent time management, and proposed the design scheme of how devices can be connected together in an accurate time framework to fit the device with the maximum sampling rate. Since not all equipment have a built-in sync tool, this problem gives rise to designing a system, may it be hardware or software, that will make data synchronization possible. In this article, three devices are used in obtaining data through simultaneous recording and individual device information is discussed below.

3. Materials and Methods.

3.1 Equipment Information. The equipment used in recording EEG data are g.tec USBAmplifier (USBamp) and g.tec GAMMAbox with a capacity of 16 channels and sampling frequency at 512 Hz. The software used in EEG recording is MATLAB R2014a. The eye-tracker device is Tobii Pro Glasses 2 together with Tobii Pro Glasses 2 Prescription Lenses for people who have imperfect vision. Tobii Pro Glasses Controller

software is used in exporting eye-tracker movie data and further eye-tracker data analysis is done through a created MATLAB program. The Perception Neuron 2.0 equipment is used in tracking motion and data is analyzed using the Axis Neuron Software.

3.2 Trigger Signal Information. To synchronize the data obtained from three devices, it is necessary for each one to have a trigger signal. EEG and Eye-tracker devices already have a built-in sync tool, while the motion capture system was not designed to send an accurate trigger signal to a high-sampling rate system.

3.2.1. Trigger Signal for EEG Data. The EEG trigger signal is dependent on the EEG experiment. Trigger behavior can be set per experiment, depending on the occurrence of an event to be observed. In Figure 2, the EEG trigger signal is generated by a photoelectric sensor. The pulse changes whether light or dark-colored surface is shown in the visual stimulus. In this case, the start of the EEG experiment was marked by the long pulse in the beginning and the following pulses mark the presentation of stimuli.




FIGURE 2. Trigger signal of EEG

3.2.2. Trigger Signal for Eye-tracker Data. The eye-tracking device has a built-in trigger for timing. As seen in Figure 3, The trigger signal has a cue for the start of eye-tracker data recording, which is the first three pulses with the frequency of 1 Hz. After that, a pulse is generated every 10 seconds. This behavior of the trigger signal is constant in every eye-tracker recording.




FIGURE 3. Trigger signal of eye-tracker

3.2.3. Trigger Signal for Motion Capture System Data. The motion capture system device does not have a built-in trigger to determine the timing, so a circuit that outputs trigger signals for the motion capture system data was created. The time series is from a transistor-transistor logic (TTL) signal output from the created trigger circuit, which can mark the start and end of recording of the motion capture system software, as well as specific events since the created circuit can either output (1) single pulse, or a (2) periodic pulse. In the case seen in Figure 4, the trigger signal for the motion capture system is periodic where one pulse is 100ms long. This gives a frequency of 5 Hz for the motion capture system trigger signal.




FIGURE 4. Trigger signal of Motion Capture System

3.3. Connection of the trigger signals. For the trigger signals to be recorded, the trigger signal circuits of (1) EEG, (2) eye-tracker, and (3) motion capture system should be connected to the digital input of the EEG equipment (g.tec USBamp), being the one having a high-sampling rate system among the other devices. The cable for digital input of the g.tec USBamp is split into ports of different voltages such as 1V, 2V, 4V, and 8V, which means the trigger signal will have the amplitude value of either 1V, 2V, 4V, and 8V,

depending on which port it is connected to. In this article, the amplitudes of trigger signals chosen and assigned for each device are seen in Table 1. This approach makes use of the amplitude as the clear differentiating factor between the three trigger signals, which is useful for signal discrimination later on.

TABLE 1. Trigger Signal Voltage Assignments for Each Device

Device	Amplitude (V)
EEG	8
Eye-tracker	2
Motion Capture System	4

3.4. Multi-trigger Recording. All trigger signals are sent simultaneously to the digital input channel, the last channel in the EEG data recording, to assure the timing accuracy due to the EEG system having the maximum sampling rate. In the set-up, there are three different trigger signals being sent to only one digital input channel. In the event that two or more signals occur at the same time, a sum of the signals will be obtained. There is an observable shift in amplitude due to this, but the values are still consistent with the individual amplitudes of the three trigger signals as seen in Table 1.




FIGURE 5. The last channel in an EEG recording dedicated for digital inputs.

4. Results and Discussion. Simultaneous recording of the three devices makes the trigger signals successfully send to the last channel in the EEG recording designated for digital input, as seen in Figure 5. The trigger signal recorded as seen in Figure 5, was used as an input into a created MATLAB program for finding each peak of the graph. Results are shown in Table 2, where the time increment for each pulse occurrence and the corresponding amplitude were obtained.

TABLE 2. Recorded trigger channel signal sample data

Time (s)	Amplitude (V)
54.142578125	14
54.144531250	14
54.146484375	12

4.1. Trigger Signal Discrimination. Data from Table 2 is then used for amplitude discrimination to distinguish which device the trigger signal is coming from. A MATLAB program was again created and used for the function of separating the trigger signals generated from different devices, from the recorded sum of trigger signals data that came from the last channel of the EEG recording. A sample of a trigger signal reconstruction data for each device is provided in Table 3.

TABLE 3. Sample Trigger Signal Reconstruction Data for Each Device

EEG		Eye-tracker		Motion Capture System	
Time (s)	Amp. (V)	Time (s)	Amp. (V)	Time (s)	Amp. (V)
50.972656250	0	63.175781250	0	50.138671875	0
50.974609375	8	63.177734375	2	50.140625000	4
50.976562500	8	63.179687500	2	50.142578125	4

Figures 6B, 6C, and 6D, show the result of each device's trigger signal reconstruction

from Figure 5, which is now illustrated more clearly as Figure 6A. Tracing lines are put to indicate which device the signal is from. Even the slightest change is accurately recovered in the signal reconstruction. Changes in the trigger signal happen quickly and is the reason why 11 significant figures are used in the time increments. This enables accurate data analysis since occurrence of events can be traced down to the nanosecond in the recorded trigger signal and it can be determined which device was sending a trigger signal at a specific time.




FIGURE 6. (A) Trigger signals of the three devices generated at the same time, (B) Trigger signal reconstruction for eye-tracker, (C) Trigger signal reconstruction for motion capture system, and (D) Trigger signal reconstruction for EEG.

4.2. Data Synchronization. The three trigger signals can then be synchronized by finding the common starting time. From here, the corresponding data of each device can also be synchronized for data matching and analysis, as seen in Figure 7.




FIGURE 7. Synchronized start time for the trigger signal of the three devices

This methodology of human EEG data analysis involving time series of gaze and motion enables the experimenters to simultaneously record EEG, eye-tracker, and motion capture system in order to provide a significant data analysis.

5. Applications. This method was partly introduced in the simultaneous analysis of EEG with hand movements in the interactive game [11] which benefits EEG experiments involving visuomotor tasks to exhibit a specific timing coordination when handshaking rhythms are synchronized together in two persons during the rock–paper–scissors game. Experiments with the proposed method can extend to the analysis of various face-to-face interactions concerned with bodily motion, and motor skill rehabilitation, which is highly affected by the visuomotor processing of the brain.

6. Extended Experimental Scheme. This innovative and consistent approach could be applied in EEG experiments to understand how humans and robots synchronize their own movements, as currently investigated by Jouaiti et al. [15]. They hypothesized that a computational model of a central pattern generator (CPG) based on non-linear relaxation oscillators can express the natural coordination mechanism not only in a coupling of humans, but also in the interaction between a human and a robot embedded with CPG inside. Indeed, they successfully demonstrated the reproduction of the same coupling phenomenon with the robot or a human. They used currently a consistent experimental setup, illustrated in Figure 1 including a TMSI Refa amplifier for EEG (32 channels, 512 Hz), T-Sens motion sensors [16] and the Open Pose system [17] for the human hand motion estimation as inputs for the model (CPG consists of Rowat-Selverston neurons [18] calculated by the Runge-Kutta numerical integration). Instead of the circuit described in Section 3.2.3, a Nano Arduino card controlled by OpenVIBE scenarios [19] was introduced to send the triggers to the T-Sens. This hardware synchronization method could be replaced and compared to the proposed method.

7. Conclusions. In the present study, an effective technique for integrating and data synchronization of EEG, gaze, and motion data involved in the simultaneous recording is presented. Each device had a trigger signal that was used for synchronizing multiple data obtained in the experiment. Trigger signals from three different devices are sent and recorded together with the EEG data. The exact time of signal occurrence of the individual trigger are determined through amplitude discrimination. Moreover, individual signal reconstruction was accomplished and was highly accurate with respect to the original data. This methodology will contribute in providing specific event analysis based on data from multiple devices and in realizing the reproducible experimental scheme where each element is replaceable in the future. Indeed, the scheme is mostly applicable in various experiments including EEG measurements and offer a wide possibility in device developments for rehabilitation and communications.

Acknowledgment. This work was supported in part by JSPS KAKENHI (15H05874, 16H01616, 17H06383) and the New Energy and Industrial Technology Development Organization (NEDO).

REFERENCES

- [1] M. Baker, Over half of psychology studies fail reproducibility test, *Nature*, News, Aug 27, 2015 doi:10.1038/nature.2015.18248.
- [2] S. Ahn, T. Nguyen, H. Jang, J. Kim and S. Jun, Exploring Neuro-Physiological Correlates of Drivers' Mental Fatigue Caused by Sleep Deprivation Using Simultaneous EEG, ECG, and fNIRS Data, *Frontiers in Human Neuroscience*, vol. 10, Article 219, 2016.
- [3] W. Goebel and C. Palmer, Temporal Control and Hand Movement Efficiency in Skilled Music Performance, *PLoS ONE*, vol.8, no.1, e50901, 2013.
- [4] J. Muraskin, J. Sherwin and P. Sajda, Knowing when not to swing: EEG evidence that enhanced perception–action coupling underlies baseball batter expertise, *NeuroImage*, vol.123, pp.1-10, 2015.
- [5] T. Thompson, T. Steffert, T. Ros, J. Leach and J. Gruzelier, EEG applications for sport and performance, *Methods*, vol.45, no.4, pp.279-288, 2008.
- [6] Peterson and D. Ferris, Differentiation in Theta and Beta Electro cortical Activity between Visual and Physical Perturbations to Walking and Standing Balance, *eneuro*, vol.5, no.4, pp.1-20, 2018.
- [7] B. Nouredin, P. Lawrence and G. Birch, Online Removal of Eye Movement and Blink EEG Artifacts Using a High-Speed Eye Tracker, *IEEE Transactions on Biomedical Engineering*, vol. 59, no.8, pp.2103-2110, 2012.
- [8] M. Plöchl, J. Ossandón and P. König, Combining EEG and eye tracking: identification, characterization, and correction of eye movement artifacts in electroencephalographic data, *Frontiers in Human Neuroscience*, vol. 6, Article 278, 2012.
- [9] G. Ai, N. Sato, B. Singh, H. Wagatsuma, Direction and viewing area-sensitive influence of EOG artifacts revealed in the EEG topographic pattern analysis, *Cognitive Neurodynamics*, vol. 10, no. 4, pp. 301–314, 2016.
- [10] B. Singh, H. Wagatsuma, A Removal of Eye Movement and Blink Artifacts from EEG Data Using Morphological Component Analysis, *Computational and Mathematical Methods in Medicine*, vol. 2017, Article ID 1861645, 2017.
- [11] G. Ai, M. Hagio, M. Ichiki, H. Wagatsuma, Simultaneous Analysis of EEGs and Movements in Interactive Hand Shaking Required Skills to Synchronize Cooperatively in Game, *2018 IEEE International Conference on Systems, Man, and Cybernetics (SMC)*, pp.590-594, 2018.
- [12] A. Delorme, T. Mullen, C. Kothe, Z. A. Acar, N. Bigdely-Shamlo, A. Vankov, and S. Makeig, EEGLAB, SIFT, NFT, BCILAB, and ERICA: New Tools for Advanced EEG Processing, *Computational Intelligence and Neuroscience*, vol. 2011, Article ID 130714, 2011.
- [13] J.-M. López-Gil, J. Virgili-Gomá, R. Gil, T. Guílera, I. Batalla, J. Soler-González, and R. García, Method for Improving EEG Based Emotion Recognition by Combining It with Synchronized Biometric and Eye Tracking Technologies in a Non-invasive and Low Cost Way, *Frontiers in Computational Neuroscience*, vol. 10, Article 85, 2016.
- [14] P. M. R. Reis, F. Hebenstreit, F. Gabsteiger, V. von Tschärner, and M. Lochmann, Methodological aspects of EEG and body dynamics measurements during motion, *Frontiers in Human Neuroscience*, vol. 8, Article 156, 2014.
- [15] M. Jouaiti, P. Hénaff, Motor Coordination Learning for Rhythmic Movements, *2019 Joint IEEE 9th International Conference on Development and Learning and Epigenetic Robotics (ICDL-EpiRob)*, 2019.
- [16] TEA CAPTIV T-Sens Respiration, <https://teaergo.com/wp/product/t-sens-respiration-tea/>.
- [17] Z. Cao, T. Simon, S.-E. Wei, Y. Sheikh, Realtime Multi-Person 2D Pose Estimation using Part Affinity Fields, *Computer Vision and Pattern Recognition (CVPR 2017)*, <https://arxiv.org/abs/1812.08008>, 2017.
- [18] M. Jouaiti, L. Caron and P. Henaff, Hebbian Plasticity in CPG Controllers Facilitates Self-Synchronization for Human-Robot Handshaking, *Frontiers in Neurorobotics*, vol. 12, 2018.
- [19] Y. Renard, F. Lotte, G. Gibert, M. Congedo, E. Maby, V. Delannoy, O. Bertrand, A. Lécuyer, OpenViBE: An Open-Source Software Platform to Design, Test and Use Brain-Computer Interfaces in Real and Virtual Environments, *Teleoperators and Virtual Environments*, Massachusetts Institute of Technology Press (MIT Press), vol. 19, no. 1, pp.35-53, 2010.