



HAL
open science

A new vessel-based method to estimate automatically the position of the non-functional fovea on altered retinography from maculopathies

Aurelie Calabrese, Vincent Fournet, Severine Dours, Frédéric Matonti, Eric Castet, Pierre Kornprobst

► To cite this version:

Aurelie Calabrese, Vincent Fournet, Severine Dours, Frédéric Matonti, Eric Castet, et al.. A new vessel-based method to estimate automatically the position of the non-functional fovea on altered retinography from maculopathies. ARVO 2022 - Annual Meeting Association for Research in Vision and Ophthalmology, May 2022, Denver, United States. hal-03600375

HAL Id: hal-03600375

<https://inria.hal.science/hal-03600375>

Submitted on 7 Mar 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Title

A new vessel-based method to estimate automatically the position of the non-functional fovea on altered retinography from maculopathies.

Authors

Calabrèse Aurélie¹, Vincent Fournet², Severine Dours¹, Frédéric Matonti³, Castet Eric¹, Kornprobst Pierre²

Affiliations

1 Aix-Marseille University, CNRS, Marseille, France

2 Université Côte d'Azur, Inria, France

3 Centre Paradis Monticelli, Marseille, France

Purpose:

In pathological fundus images with maculopathies, the fovea position is usually located using Normative Anatomical Measures (NAM). This simple method relies on two conditions: that images are acquired under standard testing conditions (primary head position and central fixation) and that the optic disk is entirely visible on the image. However, these two conditions are not always met in the case of maculopathies, especially during fixation tasks. Here, we propose a new Vessel-Based Fovea Localization (VBFL) approach.

Methods:

The spatial relationship between fovea location and vessel characteristics (density and direction) is learned from 840 annotated healthy fundus images and then used to predict the precise fovea location in new images. We evaluate our method on three different categories of fundus images: healthy (198 images from 21 eyes, each acquired with the combination of five different head positions and two fixation locations), healthy with simulated lesions (3564 images, with lesions ranging from 20 to 400 deg²) and 89 pathological fundus images collected in AMD patients.

Results:

For normal images taken under standard conditions, the NAM method yields a mean fovea localization error of 1.03° of visual angle (95% CI [0.80, 1.33]), which is slightly, but not significantly reduced by 37% with VBFL. For all conditions where the head is tilted to the side, the NAM estimation error is significantly multiplied by 4.42° (95% CI [3.45, 5.64], $p < 0.001$) while the VBFL method yields no significant increase in error prediction, with an average value of 1.21° (95% CI [0.93, 1.56]), representing a 73% reduction in prediction error compared to the NAM method. With simulated lesions, performance of the VBFL method decreases significantly as lesion size increases, with a mean error ranging from 0.83° at 20 deg² up to 2.3° at 400 deg². Overall, performance of VBFL remains better than NAM until lesion size reaches 200 deg². For pathological images, the error distribution with VBFL is not higher than for healthy data, suggesting that actual AMD lesions do not negatively affect the method's performance.

Conclusions:

The vascular structure provides enough information to precisely locate the fovea in fundus images in a way that is robust to head tilt, eccentric fixation location, missing vessels, and real macular lesions.