



HAL
open science

Multi-contact epineural electrical stimulation to restore upper-limb functions

Christine Azevedo Coste, Charles Fattal, Antoine Geffrier, Lucas Oliveira Fonseca, Lucie William, Arthur Hiairassary, David Andreu, Jacques Teissier, David Guiraud

► To cite this version:

Christine Azevedo Coste, Charles Fattal, Antoine Geffrier, Lucas Oliveira Fonseca, Lucie William, et al.. Multi-contact epineural electrical stimulation to restore upper-limb functions. RehabWeek 2022, Jul 2022, ROTTERDAM, Netherlands. hal-03702524

HAL Id: hal-03702524

<https://inria.hal.science/hal-03702524v1>

Submitted on 2 Jan 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Multi-contact epineural electrical stimulation to restore upper-limb functions

Christine Azevedo Coste¹, Charles Fattal^{1,2}, Antoine Geffrier^{4,6}, Lucas Fonseca¹, Lucie William¹, Arthur Hiairassary^{1,5}, David Andreu⁵, Jacques Teissier³, David Guiraud^{1,5}

¹INRIA, University of Montpellier, Montpellier, France -

²Rehabilitation Center Bouffard-Vercelli USSAP, Perpignan, France

³Surgical Clinic OrthoSud, Montpellier, France

⁴University Hospital Center Pontchaillou, Rennes, France

⁵Neurinnov, Montpellier, France, France

⁶APHP, Paris France

Christine.Azevedo@inria.fr

Abstract: We investigated the feasibility of a novel approach aiming at restoring functional movements in completely paralyzed upper limbs using neural stimulation. Two multi-contact cuff electrodes were wrapped around radial and median nerves, unilaterally, in 2 individuals with complete high tetraplegia. Electrodes were maintained for 28 days. Advanced stimulation configurations were implemented to achieve selective activation of fascicles and elicit different grasping. A control interface was developed to allow users to pilot movements of their hand by triggering stimulation. Both participants were able to execute palmar and key pinch grasping during functional tasks.

Keywords: tetraplegia, FES, cuff neural electrodes, selectivity.

Introduction

Functional electrical stimulation (FES) makes possible the restoration of upper limb function in individuals with complete tetraplegia when functional tendon surgery or nerve transfers are not possible [1]. The majority of approaches that investigated implanted electrical stimulation used intramuscular or epimysial electrodes [2,3]. This implies implanting as many electrodes and wires as the number of muscles to be stimulated. Instead, we investigated the possibility of stimulating nerves proximally to reduce the number of implanted electrodes. In a previous study, we tested multi-source epineural stimulation of the median and the radial nerves intra-operatively in 8 patients with complete tetraplegia [4]. For all subjects, under general anaesthesia, it was possible to selectively stimulate muscle groups to generate various movements depending on the current path over the multi-contact electrodes. Based on this, we aimed to study the feasibility of the approach during a 28-days period with awake users [7]. The objectives were to study safety of procedures and feasibility of producing functional movements autonomously controlled by the users by means of an interface taking advantage of the remaining voluntary activities of contralateral upper limb. Some information is provided in the following and will be developed during presentation at the conference.

Methods

2 subjects with high tetraplegia C4, AIS A, group 0 of the International Classification for Surgery of the Hand in Tetraplegia were included. They signed informed consent. The study was approved by the National Ethics Committee (ClinicalTrials.gov identification number NCT04306328). The electrodes were placed during a first surgical procedure at day 0 and the explantation was performed 28 days later during a second surgical procedure. The full device was composed of 2 multi-contact cuff electrodes (CorTec GmbH Freiburg Germany, 8 central contacts for the median, 6 central contacts for the radial) with their individual percutaneous cable and extracorporeal connector, 2 extracorporeal cables, 1 external multi-source stimulator (STIMEP, INRIA, Montpellier) [8] connected to a computer (figure 1). Wireless sensors (Delsys, Natick, MA) were placed on the skin to measure voluntary muscle signals (electromyography (EMG)) and voluntary movements (inertial measurement units (IMU)). These signals were processed to detect the user piloting orders.




Figure 1: Principle of the solution proposed to restore hand function in individuals with complete tetraplegia. 2 multi-contact epineural electrodes are wrapped around median and radial nerves just above elbow.

4 to 8 EMG sensors (Delsys, Natick, MA) were also placed on the skin to wirelessly collect: 1) evoked electromyography (eEMG) data from muscle contractions

induced by FES, 2) forces applied on instrumented objects (force resistive sensors (FSR)). The eEMGs were post-processed to extract recruitment curves [9] and FSR data used to assess the pinch or grasping forces. 2 control modalities were explored [6] for the user to trigger the stimulation autonomously using the contralateral limb: 1) supraslesional voluntary contractions captured by EMG and detected by a threshold detection algorithm, 2) voluntary stereotyped shoulder movements captured by IMU and processed by a classifier algorithm to detect predefined shoulder movements. A finite state machine (FSM) was defined to associate user commands (EMG threshold detection or recognition of a predefined movement (IMU)) with actions depending on the current FSM state. Participants performed 3 sessions a week over the 28 days when the parameters were tuned and optimised and induced movements were assessed.

Results

Tests confirmed the ability to produce multiple movements including functional and reproducible key-grip and digito-palmar grasping. In participant 1, voluntary muscle contraction detection and voluntary movement detection control interface modalities were successful. For participant 2, only the voluntary movement detection modality was possible due to an important fatigability of the shoulder muscles. The success rate in task execution was estimated using Motor Capacity Scale C score at 54% for patient 1 and 51% for patient 2. The results will be detailed during the presentation at the conference.



Figure 2: Example of 3 tasks achieved by participant 1. Top: key grip pinch. Middle and Bottom: palmar grasp.

Discussion

Our study suggests that multi-contact cuff neural combined with multi source electrical stimulation offers the possibility to selectively activate nerve fascicles in a reproducible manner. Multipolar neural electrical stimulation of the median nerve and radial nerve allowed

for useful grasps and release in both participants. Future studies will consolidate this proof of concept by extending the number of participants.

Acknowledgement

Authors would like to acknowledge the participants, Mrs Caralp and Porra (OT), Mrs Panel and Ferrer (PT), as well as the health care teams in charge of day-to-day care. The study was supported by grants from EIT Health, Region Occitanie and ERC MSCA.

References

- [1] van Zyl, N., Hill, B., Cooper, C., Hahn, J., and Galea, M.P. (2019). Expanding traditional tendon-based techniques with nerve transfers for the restoration of upper limb function in tetraplegia: a prospective case series. *Lancet* 394, 565–575.
- [2] Hoshimiya, N., and Handa, Y. (1989). A master-slave type multi-channel functional electrical stimulation (FES) system for the control of the paralyzed upper extremities. *Automedica*, 209–220.
- [3] Keith, M.W., Peckham, P.H., Thrope, G.B., Stroh, K.C., Smith, B., Buckett, J.R., Kilgore, K.L., and Jatich, J.W. (1989). Implantable functional neuromuscular stimulation in the tetraplegic hand. *J Hand Surg Am* 14, 524–530.
- [4] Tigrá, W., Dali, M., William, L., Fattal, C., Gélis, A., Divoux, J.-L., Coulet, B., Teissier, J., Guiraud, D., and Azevedo Coste, C. (2020). Selective neural electrical stimulation restores hand and forearm movements in individuals with complete tetraplegia. *J Neuroeng Rehabil* 17, 66.
- [5] Tigrá, W., Navarro, B., Cherubini, A., Gorrón, X., Gelis, A., Fattal, C., Guiraud, D., and Azevedo Coste, C. (2018). A Novel EMG Interface for Individuals with Tetraplegia to Pilot Robot Hand Grasping. *IEEE Trans Neural Syst Rehabil Eng* 26, 291–298.
- [6] Fonseca, L., Tigrá, W., Navarro, B., Guiraud, D., Fattal, C., Bó, A., Fachin-Martins, E., Leynaert, V., Gélis, A., and Azevedo-Coste, C. (2019). Assisted Grasping in Individuals with Tetraplegia: Improving Control through Residual Muscle Contraction and Movement. *Sensors (Basel)* 19.
- [7] Fattal C., Teissier J., Geffrier A., Fonseca L., William L., Andreu D., Guiraud D., Azevedo-Coste C (2022). Restoring Hand Functions in People with Tetraplegia through Multi-Contact, Fascicular, and Auto-Pilot Stimulation: A Proof-of-Concept Demonstration. *Journal of Neurotrauma*.
- [8] Guiho T., López-Alvarez V., Cvancara P., Hiairassary A., Andreu D., Stieglitz T., Navarro X., Guiraud D. (2021), New Stimulation Device to Drive Multiple Transverse Intrafascicular Electrodes and Achieve Highly Selective and Rich Neural Responses, *Sensors*.
- [9] William L., Dali M., Azevedo Coste C., Guiraud D. (2022) A method based on wavelets to analyse overlapped and dependant M-Waves, *Journal of Electromyography and Kinesiology*.