



HAL
open science

Standardised FES-induced fatigue-testing of paralysed human quadriceps muscles during a dynamic movement task

Martin Schmoll, Ronan Le Guillou, David Lobato Borges, Charles Fattal, Emerson Fachin-Martins, Christine Azevedo Coste

► To cite this version:

Martin Schmoll, Ronan Le Guillou, David Lobato Borges, Charles Fattal, Emerson Fachin-Martins, et al.. Standardised FES-induced fatigue-testing of paralysed human quadriceps muscles during a dynamic movement task. FES Workshop 2022 - International Workshop on Functional Electrical Stimulation, Sep 2022, Innsbruck, Austria. hal-03924013

HAL Id: hal-03924013

<https://inria.hal.science/hal-03924013>

Submitted on 5 Jan 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Standardised FES-induced fatigue-testing of paralysed human quadriceps muscles during a dynamic movement task

Martin Schmoll^{1,2}, Ronan Le Guillou¹, David Lobato Borges³, Charles Fattal⁴, Emerson Fachin-Martins³, Christine Azevedo-Coste¹

¹INRIA - LIRMM, Université de Montpellier, Montpellier, France

²Center for Medical Physics and Biomedical Engineering, Medical University of Vienna, Vienna, Austria

³NTAAI, Faculdade de Ceilândia, Universidade de Brasília, Brasília, Brazil

⁴Rehabilitation Center Bouffard Vercelli, USSAP, Perpignan, France

Introduction

Muscular fatigue remains a persistent problem during functional electrical stimulation (FES). One technique known as spatially distributed sequential stimulation (SDSS), uses multiple electrodes to stimulate distinct pools of motor-units at a lower frequency to decrease fatigue. Although literature claims clear benefits, interpretation of the results and applicability proves difficult due to heterogeneous testing strategies. Thus, we present a standardized method for fatigue testing in individuals with spinal cord injury (SCI), tailored to the requirements of a practical application (e.g. FES-Cycling).

Methods

The fatigue-development of 6 paralysed quadriceps muscles of 3 different participants with complete paraplegia was assessed via a fatigue-index during 180 dynamic contractions, comparing different SDSS electrode-configurations against conventional single-channel stimulation. For standardisation, testing was conducted at 40% of the peak-torque during a maximal evoked contraction (MEC), in previously trained individuals with SCI.

Results

Our results were unable to detect a significant difference in the fatigue-development comparing SDSS against conventional stimulation. Indifferent results regarding the potential benefits of SDSS in a practical setting are widely reported across various international research groups. We hypothesise that the positive effects of SDSS diminish with increasing stimulation amplitudes (required to elicit strong contractions), due to a loss of selectivity.

Conclusion

Assessing muscular fatigue is not a trivial task. To allow for a successful practical translation of a technique, it is crucial to perform fatigue-testing tailored to the requirements of a potential use-case. In practical applications, FES is often used to elicit forceful dynamic contraction in individuals with SCI who already conducted a dedicated FES-training program. Therefore, we recommend to assess fatigue-development at higher forces (e.g. 40% MEC) in pre-trained individuals with SCI to better reflect the practical demands of FES-applications.