

Standardized numerical simulations of cardiac electrical stimulation devices

Valentin Pannetier, Yves Coudière, Michael Leguèbe

► To cite this version:

Valentin Pannetier, Yves Coudière, Michael Leguèbe. Standardized numerical simulations of cardiac electrical stimulation devices. Journée annuelle de l'EDMI, Apr 2023, Bordeaux (France), France. hal-04384583

HAL Id: hal-04384583 https://inria.hal.science/hal-04384583

Submitted on 10 Jan2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

Standardized numerical simulations of cardiac electrical stimulation devices

ODE



(mV)

Amplitude

 (\mathbf{X})

No capture

Valentin Pannetier, Yves Coudière, Michael Leguèbe Université de Bordeaux, IMB, Inria and IHU Liryc

CONTEXT

The **SimCardioTest** project aims to improve the interest and benefits of **insilico simulations** in the case of cardiovascular diseases.

We work on the electrical part with **MicroPort CRM pacemakers** manufacturers.

An **interface** separates the electrode from the tissue, where several **biochemical** effects occur. We want to understand the impact of contact **impedance** during a stimulation and quantify the **energy** actually applied to the tissue.

STIMULATIONS

Parametrized by their **amplitude** and **duration** of pulse.

Efficient stimulations lead to a capture, and the other ones fail.

 The plane of stimulations is splitted in two parts by the Lapicque curve

 $A(d) = A_{\rm rh} \left(1 + \frac{t_{\rm ch}}{d} \right)$

with two parameters called rheobase and chronaxie.

PDE – 3D MODEL

 \checkmark


Duration (ms)

Capture

ODE – OD MODEL

Why

- Basic study on electrical properties of **contacts**.
- Test some algorithms, and simplified **nonlinear** effects.
- First comparison with **experimental** data.



Device equation (equivalent RC circuit)

$\mathrm{d}I$		_ 1	(V_1)	V_2	$I_{ m ion}$	$V_{\rm m}$
$\overline{\mathrm{d}t}$	τ – – –	\overline{R}	$\left(\frac{-}{\tau_1}\right)$	$+ - + \tau_2$	$\overline{C_{\rm m}}$	$+ \frac{V_{\mathrm{m}}}{\tau_{\mathrm{m}}} ight)$

Contact equations (tip and ring)

 $\frac{\mathrm{d}V_i}{\mathrm{d}t} + \frac{V_i}{\tau_i} = \frac{I}{C_i}$

* **OD cell** equation with **ionic terms** from literature

$$\frac{\mathrm{d}V_{\mathrm{m}}}{\mathrm{d}t} + \frac{I_{\mathrm{ion}}}{C_{\mathrm{m}}} + \frac{V_{\mathrm{m}}}{\tau_{\mathrm{m}}} = \frac{I}{SC_{\mathrm{m}}}$$

 $\frac{\mathrm{d}h}{\mathrm{d}t} + g(v_{\mathrm{m}}, h) = 0$

How to choose and determine contact and scaling parameters...?

Wнy

- Take geometry into account (electrode distances, front wave propagation...).
- Compare **experimental** data with our solution.

Bidomain equations in the tissue


- $-\operatorname{div}\left(\sigma_{\mathrm{i}}\nabla u_{\mathrm{i}}\right) = -C_{\mathrm{m}}\partial_{t}v_{\mathrm{m}} j_{\mathrm{ion}}$ $-\operatorname{div}\left(\sigma_{\mathrm{e}}\nabla u_{\mathrm{e}}\right) = C_{\mathrm{m}}\partial_{t}v_{\mathrm{m}} + j_{\mathrm{ion}}$ $\partial_{t}h + g(v_{\mathrm{m}}, h) = 0$
- Poisson equation in the blood
- $-\mathrm{div}\left(\sigma_{\mathrm{e}}\nabla u_{\mathrm{e}}\right) = 0$

PDE

- * At **boundaries** with **electrodes** (tip and ring)
- $-\sigma_{\rm e}\nabla u_{\rm e}\cdot\mathbf{n}=j_-$
- $-\sigma_{\rm e}\nabla u_{\rm e}\cdot\mathbf{n}=j_+$
- * At **exterior boundaries**, a transmissive b.c. $-\sigma_{\rm e} \nabla u_{\rm e} \cdot \mathbf{n} = \mu_{\rm e} u_{\rm e}$ $-\sigma_{\rm i} \nabla u_{\rm i} \cdot \mathbf{n} = \mu_{\rm i} u_{\rm i}$


Proof of existence on the way !

How to adjust conductivities, ionic model, all parameters...?



RINGER TESTS

Measurements in saline solution



MicroPort


CONTEXT

These experiments were done in a laboratory of the industrial partner company **MicroPort**.

PRINCIPLE

A lead is **submersed** in a salt bath, **voltage** and **current** passing through between the two electrodes are measured.

SHAPE OF DATA OBTAINED



DISCHARGE Phase to counterbalance capacitive effects of contacts.

CONTEXT


These experiments were done at **IHU Liryc** with MicroPort leads.

PRINCIPLE

A **sheep**'s ventricle is placed in a bath with two **cameras** pointing at it (endocardial and epicardial sides). A lead is then **implanted** and **stimulates** the tissue. An **activation map** is calculated from the fluorescence captured by the cameras.

EX-VIVO DATA FROM 2 ANIMALS

Activation map by optical mapping.
Measured voltage at lead terminal.
Geometries and some tissue characteristics by MRI 9.4T.



OPTICAL MAPPING



RESULTS

CONTACT CALIBRATION


Several contact models tested (different ODE systems).
We are confronted with the non identifiability of

NO RESULTS IN 3D FOR LAPICQUE STUDIES YET

But some works on bidomain coupled with an extracardiac



RESULTS



parameters.

* Still a large variability even with R//C.

Accurate enough to reproduce measures in saline solution.

SCALING PARAMETERS

 \ast Great influence of parameter S (surface membrane). \ast S and τ_m are **surrogates** of the geometry.

LAPICQUE CURVES

Obtained by 2D dichotomy for now.

* Not accurate enough for **ex-vivo comparison**.

Order of magnitude and results relevant for exploration.

domain already done.


OBJECTIVES

- * Reproduce 0D studies in 3D.
- Quantify influence of tissue variability and geometric properties on results.

For that we need to exploit experimental data, and several **calibrations** will be required.

ALL THIS WORK TO

Take into account **pacemaker** in models.
Create a web platform with a design tool for lead manufacturers.







This project received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No 101016496